

LightMix® OneTube Dental Bac I (RUO)

Qualitative multiplex PCR detection of dental bacterial targets for research use only testing on LightCycler® instruments.



Convenience



Consistency



Fewer plastics and less waste

Primer and probe sets have been premixed and combined for you, reducing error risks associated with complex mastermix preparation. Additionally, you'll receive fewer plastics creating less waste which can support your sustainability goals.

Streamline your workflow with pre-designed, pre-mixed multiplex primer/probe sets, courtesy of TIB Molbiol, for use on LightCycler® 480 platforms¹.



Product concept

1x vial with 5 sets of primer/probes for 96 reactions/20 µL volume.
1x vial Positive Control



Inclusivity

Qualitative multiplex PCR detection and discrimination of 5x dental bacterial targets:
1) *Actinobacillus actinomycetemcomitans*, **2)** *Porphyromonas gingivalis*, **3)** *Prevotella intermedia*, **4)** *Tannerella forsythia*, **5)** *Treponema denticola*



Exclusivity

No other oral flora will be detected



Sensitivity

This assay detects 10 genome equivalent copies or less per reaction (plasmid DNA dilution).
LoD not tested



Sampling

Typical specimens are saliva or gingival crevicular fluid.



Usage

Refer to **Instructions for Use**. Default thermal cycling protocol recommended.



Labelling

RUO²



Stability

18 months

¹Compatible with LightCycler® 480 & 480 II²

²For Research Use Only. Not for use in diagnostic procedures.

Reference: LightMix® OneTube Dental Bac I, Package Insert

Ordering Information

Product name	Reactions/volume	Material number
LightMix® OneTube Dental Bac I (RUO)	96 reactions/20 µL	10409312001
LightCycler® Multiplex DNA Master	200 reactions/20 µL	07339585001
	1000 reactions/20 µL	07339577001
PhHV Extraction Control (660)	96 reactions/20 µL	07093802001

For Research Use Only. Not for use in clinical diagnosis.

Channel configuration

Five optical channels are used to detect the following targets. The 660 channel is recommended for the PhHV Extraction Control.

500	530	580	610	640	660
A. actinomyce- temcomitans	P. gingivalis	P. intermedia	T. forsythia	T. denticola	PhHV EC



LightCycler® 480 II System

Reference: LightMix® OneTube Dental Bac I, Package Insert

LightMix® is a trademark of TIB Molbiol.
LIGHTCYCLER is a trademark of Roche.

© 2025
All rights reserved.

Published by

Roche Diagnostics
9115 Hague Road
Indianapolis, IN 46250

MC-US-17307

diagnostics.roche.com